



Summary of Safety and Clinical Performance (SSCP) Template

<https://clinicalstudytemplates.com/>

Template Use

This Summary of Safety and Clinical Performance (SSCP) Template is specifically designed to be used in conjunction with the Clinical Evaluation Report (CER) Template.

This SSCP Template includes all the fundamental elements of an EU MDR 2017/745-compliant SSCP per Article 32(2), and the content and sections follow the Medical Device Coordination Group (MDCG) 2019-9 guideline, conforming to the expectations by the Notified Body (NB). The structure of this SSCP enables content to be directly copied from equivalent sections of the associated CER and device IFU, with only minor modifications.

- For more information on the EU MDR 2017/745, visit <https://eur-lex.europa.eu/legal-content/EN/TXT/PDF/?uri=CELEX:32017R0745>
- For more information on MDCG guidelines, visit https://health.ec.europa.eu/medical-devices-sector/new-regulations/guidance-mdcg-endorsed-documents-and-other-guidance_en

Note that, although the EU MDR 2017/745 became effective on May 2017, the transition period for the remediation of MDD-compliant clinical evaluations to MDR-compliant clinical evaluations has been extended to December 2027 for Class III and Class IIb devices, and to December 2028 for Class IIa and Class I devices.

The integration of automated formatting features through the customized Styles gallery, and the exceptional ease of using this SSCP Template allow experienced authors to save valuable time by eliminating labor-intensive generation of standard verbiage and repetitive formatting, and offer novice writers clear guidance to create MDR-compliant, high-quality documents that ensure consistency across other SSCPs.

Taking into account the various types of clinical devices with different classifications, and Intended Purpose-associated statements, such as Intended Use, Indications for Use, Contraindications, Target Patient Population, Users, Warnings, Cautions, Warnings, Claims, and Clinical Benefits, that are all to be taken into account when developing an SSCP, authors should work with their device development team members (i.e., Research and Development, Regulatory Affairs, Clinical Affairs, Medical Affairs, Manufacturing, Post-Market Surveillance, Post-Market Clinical Follow-up, etc.) to ensure that the document's content and data are presented in a clear and concise manner, which may require this template to be modified accordingly, so that it is in line with the device's unique aspects and specific characteristics.

Text Color-coded Instructions


- **Red Text:** instructions (delete)
- **Green Text:** placeholder (replace, modify or delete; select and convert the text color to black by changing the Font Color under the Home tab)
- **Blue Text:** examples (modify or delete; select and convert the text color to black by changing the Font Color under the Home tab)
- **Black Text:** commonly-used text (modify as necessary)

Applying Styles, and Inserting Tables, Figures and Citation References

Visit <https://clinicalstudytemplates.com/tutorials/> for short videos that explain how to use the automated features of this template.

This template has been designed so that hyperlinking the headings of sections and other data (e.g., tables and figures) allows their respective Table of Contents (TOC), Table of Tables, Table of Figures (if applicable) and in-text references to data to be updated by right-clicking on the TOC, and then selecting Update Field, as long as associated data (e.g., tables, figures, citation references) are properly inserted per the instructions outlined below.

- Applying Styles:** This template contains a set of 18 pre-made styles (see the available styles of the Styles gallery displayed in the green oval below) that are quick and easy to use, and ensure uniformity among the different documents when applied correctly.




When using these styles, complete and consistent formatting is applied to the selected text with a single click. If certain words in the text need to be italicized, bolded, underlined, etc., then apply manual changes to the font only *after* the style has been applied. For the rest, follow the instructions below, and refrain from manually changing the text as much as possible.

- For general text, apply the **Body Text** style
- For headings of sections, apply the appropriate **Heading** styles (there are 5 pre-made levels)
- For numbered lists in the body text, apply the **List Numbered Level** styles (there is a first and second-level style available)
- For bulleted lists in the body text, apply the **List Bullets Level** styles (there is a first and second-level style available)
- For captions of tables and figures, the **Caption** style should be applied automatically when tagging tables and figures with clickable captions as described below
- For general text in tables, apply the **Table Body Content** style
- For numbered lists in tables, apply the **Table List Numbered Level** styles (there is a first and second-level style available)
- For bulleted lists in tables, apply the **Table List Bullets Level** styles (there is a first and second-level style available)

- There are 2 additional pre-made styles (Title without Heading and No Spacing) that can be used and modified, as necessary
- **Numbered Sections:** Headings define the start of a new section (e.g., Introduction, Study Objectives, Investigational Plan, etc.). When changing, modifying or adding a heading, make sure to tag it with the appropriate Heading style under the Home tab, so that MS Word recognizes the heading and its (sub)level as such.
To then insert a clickable link to a section (i.e., a cross reference) somewhere else in the text, place the cursor where the clickable link needs to be inserted, go to the References Tab, select Cross-reference, look under Reference type for the heading, and then insert the cross reference by choosing Heading number under the Insert reference to drop-down menu.
- **Sections without Numbers:** For subheadings that do not require a designated section number, and that should remain absent from the Table of Contents, use the Title Without Heading style.
- **Tables, Figures and other data:** To insert a table or figure that can be reached via a clickable link somewhere else in the text, first select the newly-inserted table or figure, then go to the Reference Tab, select Insert Caption; next choose the appropriate Label from the drop-down menu (i.e., Table or Figure), and click on the Numbering button to check the box for Include chapter number, if preferred. After clicking OK, a sequential number is automatically generated and assigned to the table/figure. Note that table captions appear *above* the table, and figure captions appear *below* the figure. Next, expand the caption with a descriptive text as necessary. To then reference this table or figure somewhere else in the text with a clickable link, place the cursor where the clickable link needs to be inserted, go to the Reference Tab, select Cross-reference, and look under Reference type for the referable table/figure. Note that it is best to choose Only label and number under the Insert reference to drop-down menu.
- **Citation References:** To insert a new citation reference (i.e., a reference to journal article, book, report, poster, website, etc.), first go to the Reference Tab and select Insert Citation. Then select Add new source, choose the appropriate Type of source, and fill out the required information. The reference should appear in Section 11 after updating that field. To then insert a citation somewhere in the text to reference in the reference list, place the cursor where the citation needs to be inserted, go to the Reference Tab and select Insert Citation, and then choose the appropriate reference from the presented list. To reference this reference list somewhere else in the text with a clickable link, place the cursor where the clickable link needs to be inserted, go to the Reference Tab, select Cross-reference, and look under Reference type for section 11, and then insert the cross reference by choosing Heading number under the Insert reference to drop-down menu.


- **References to Externally-linked Files:** When Appendix documents and other files (e.g., Excel sheets or PDF files) need to be referenced in the text, clickable hyperlinks can be created in the final (approved) version of the CEP, by selecting the clickable text, then right-clicking on the selection and choosing Link to hyperlink the referenced file. To be able to easily recognize these as clickable links, the font color of the clickable text can be underlined and changed to a distinguishable [blue](#).
- **Copying text from other documents:** When copying text from other documents, it is strongly advised to paste the text in this template without the source formatting, so that conflicting styles are not carried over. To do so, select and copy the source text, then right-click in this template where the source text needs to be pasted via Paste Options > Keep Text Only (T):



And then apply the desired style from the Styles gallery.

General Good Document Creation Practices

Below are some general guidelines for good document creation practices. Visualizing the paragraph formatting can be extremely helpful to display any hidden formatting and for cleaning up the document:



This template contains the most frequently used styles that are available in the Styles gallery (see the instructions above for how to apply these appropriately).

Body text can be justified, but numbered and bulleted lists as well as text in tables is generally not justified.

When a table spans multiple pages, select the descriptive top row(s) of the table, and use Repeat as header row at the top of each page under Table properties to have the descriptive top row(s) automatically reappear at the top of the table parts across multiple pages.

To start a new page, use the Page Break (short-cut key: Ctrl+Enter) to mark the end of the current page (instead of hitting the Enter key until the end of the page is reached). Note that Section Breaks should be avoided as much as possible, as they serve a different purpose (e.g., to change the page orientation from Portrait to Landscape), which is generally not needed; even if tables are large), or to change the headers/footers for a particular set of pages.

When aligning text vertically, use the appropriate Tab Stop Positions (instead of hitting the space bar or Tab keys many times).

To prevent a paragraph from being separated from a following paragraph (e.g., the text preceding a bullet list), select the paragraph, and check the *Keep with next* property under Home > Paragraph > Line and Page Breaks.

Per the University of Oxford Style Guide, there is *no* space between a number and percent, degree, or mathematical characters, like +, -, ±, >, <, ≥, ≤, ×, = (e.g., ≤5%, -4°C, and 5×3=15; be sure to use the actual degree character instead of a superscripted “o”). On the other hand, there *is* a space between a number and unit (e.g., 25 mL, 65 kg, 15 minutes); to prevent the number and unit being separated over 2 lines when at the end of the line, use the “non-breaking space” with Ctrl+Shift+Space (the same can be applied to prevent a hyphen or minus-sign getting separated from the number).

Per international standards, it is advised to use a 2-number day (dd), 3-letter capitalized month (MMM), and 4-number year (yyyy) dating format (separated by hyphens) as exemplified: August 9th, 2021 would be written as 09-AUG-2021.

Final Notes

- Delete these instructions prior to finalizing the document
- Remove the “**DRAFT**” watermark prior to finalizing the document via Design > Watermark > Remove Watermark
- Update the TOC last, so that the correct page numbers are displayed, by right-clicking on the TOC, and then selecting Update Field

– This line and everything above should be deleted prior to finalizing document. –

Summary of Safety and Clinical Performance

for

Device Name(s)

Manufactured by

Manufacturer Name

SAMPLE

Update the TOC below at the very last, and *after* the instructions pages at the beginning of this Template as well as this sentence have been deleted, to ensure that correct titles of the headings and page numbers are reflected; to update, right-click anywhere on the Table of Contents below, and then select Update Field > Update Entire Table.

TABLE OF CONTENTS

TABLE OF CONTENTS.....	8
FOR HEALTHCARE PROFESSIONALS	10
1 DEVICE IDENTIFICATION AND GENERAL INFORMATION	10
2 INTENDED USE OF THE DEVICE	11
2.1 Intended Use/Purpose	11
2.2 Indication(s) and Target Population(s)	11
2.2.1 Indications for Use.....	11
2.2.2 Target Patient Population	12
2.3 Contraindications and/or Limitations	12
3 DEVICE INFORMATION	12
3.1 Device Description	12
3.1.1 Design and Specifications.....	12
3.1.2 Principles of Operation	12
3.1.3 Device Characteristics	12
3.1.3.1 Technical Characteristics	12
3.1.3.2 Biological Characteristics.....	12
3.1.3.3 Clinical Characteristics	12
3.1.4 Device Mode of Action	12
3.1.5 Duration of Body Contact.....	13
3.1.6 In-Use Lifetime	13
3.2 Device History	13
3.3 Device Accessories.....	13
3.4 Other Devices and Products to be Used with the Device	14
4 RISKS AND WARNINGS.....	14
4.1 Residual Risks and Undesirable Effects	14
4.2 Warnings and Cautions	14
4.2.1 Warnings.....	14
4.2.2 Cautions	15
4.3 Other Safety Aspects	15
4.3.1 Corrective and Preventative Actions.....	15
4.3.2 Field Actions and Recalls.....	15
5 SUMMARY OF CLINICAL EVALUATION AND POST-MARKET CLINICAL FOLLOW-UP (PMCF)	15
5.1 Summary of Clinical Data Related to the Equivalent Device	15
5.2 Clinical Data from Investigations Conducted Before CE-Marking.....	16
5.3 Clinical Data from Other Sources.....	18
5.3.1 Post-Market Clinical Follow-up	18
5.3.2 Post-Market Surveillance.....	18

5.3.3 Scientific Literature 18

5.4 Overall Summary of Clinical Performance and Safety19

5.5 Ongoing or Planned Post-Market Clinical Follow-up20

6 POSSIBLE DIAGNOSTIC OR THERAPEUTIC ALTERNATIVES20

7 SUGGESTED PROFILE AND TRAINING FOR USERS.....20

8 REFERENCES TO HARMONISED STANDARDS AND COMMON SPECIFICATIONS.....21

9 REVISION HISTORY22

FOR PATIENTS.....23

1 DEVICE IDENTIFICATION AND GENERAL INFORMATION23

2 INTENDED USE OF THE DEVICE24

2.1 Intended Use/Purpose24

2.2 Indication(s) and Target Population(s)24

2.2.1 Indications for Use..... 24

2.2.2 Target Patient Population 24

2.3 Contraindications and/or Limitations24

3 DEVICE DESCRIPTION.....24

3.1 Device Description and Material/Substances in Contact with Patient Tissues24

3.1.1 Device Description 24

3.1.2 Material/Substances in Contact with Patient Tissues 24

3.2 Medicinal Substances in the Device24

3.3 Device Mode of Action25

3.4 Device Accessories.....25

4 RISKS AND WARNINGS.....25

4.1 How Potential Risks Have Been Controlled or Managed25

4.2 Remaining Risks and Undesirable Effects25

4.3 Warnings and Cautions26

4.3.1 Warnings..... 26

4.3.2 Cautions 26

4.4 Summary of Field Safety Corrective Actions.....26

5 SUMMARY OF CLINICAL EVALUATION AND POST-MARKET CLINICAL FOLLOW-UP (PMCF)26

5.1 Clinical Background of the Device.....26

5.2 Clinical Evidence for CE-Marking26

5.3 Safety26

6 POSSIBLE DIAGNOSTIC OR THERAPEUTIC ALTERNATIVES27

7 SUGGESTED TRAINING FOR USERS.....27